EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	ClearDetections P.O. Box 170, NL-6700 PD Wageningen, Netherlands
Short description of the test	Diagnostic Real-time PCR assay for identification and detection of Aphelenchoides besseyi
Date, reference of the validation report	2011-01-01 - 86 ; 'Validatie van moleculaire identificatie- en detectiemethoden van Aphelenchoides fragariae, A. ritzemabosi, A. subtenuis en A. besseyi'. Validation report (in Dutch) of FES study.
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	no
Description of the test	
, and the second	
Organism(s)	Aphelenchoides besseyi (APLOBE)
Detection / identification	detection and identification
Method(s)	Molecular real time PCR
Method: Molecular real time PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	yes
EPPO Diagnostic Protocol name	PM 7/039 Aphelenchoides besseyi (version 2)
Name of the test	Real-time PCR test based on SSU rDNA (Clear detection)
Is the test modified compared to the reference test	no
Kit	
Is a kit used	yes
Manufacturer name	CLEAR DETECTIONS
Specify the kit used	RT-N-D-0302 ClearDetections Real-Time PCR Diagnostic kit: Aphelenchoides besseyi
	•

Kit used following the manufacturer's instructions?	yes	
Other information		
Reaction type	Simplex	
Other details on the test	Real-time PCR; based on detection of a fluorescent DNA-binding dye	
Are the performance characteristics included in the EPPO diagnostic protocol?	yes	
Performance Criteria :		
Organism 1.:	Aphelenchoides besseyi(APLOBE)	
Analytical sensitivity		
What is smallest amount of target that can be detected reliably?	< one individual nematode (~ 3 cells of target nematode)	
<u>Diagnostic sensitivity</u>		
Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	100%	
Standard test(s)	Morphological identification	
Analytical specificity - inclusivity		
Number of strains/populations of target organisms tested	A. besseyi (2160 and E9192) obtained from Dutch PPO (ref. Gerrit Karssen)	
Specificity value	100%	
Analytical specificity - exclusivity		
Number of non-target organisms tested	Aphelenchoides subtenuis; A. fragariae; A. ritzemabosi; A. saprophilus; Ditylenchus dipsaci; D. destructor	
Specificity value	100% No cross reaction	
Diagnostic Specificity		
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	100%	
Specify the test(s)	Morphological identification	
Reproducibility		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%	
Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%	
Test performance study		
Test performance study?	no	
Other information		
Any other information considered useful	Accuracy: 100% Dynamic range: between 10-100 and 0.1 billion copies of target rDNA Selectivity:	

The following complementary files are available	www.cleardetections.com). Validation report
	100% Robustness: OK This qPCR assay for identification and detection of A. besseyi is available as all-inclusive molecular kit, including primer sets, positive control DNA, PCR enhancer and PCR mix and a bench-side protocol describing the laboratory procedure (for information visit

Creation date: 2014-11-12 00:00:00 - Last update: 2021-03-11 19:41:05