

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Council for Agricultural Research and Economics- Research Centre for Plant Protection and Certification Via Carlo Giuseppe Bertero, 22, 00156 Rome, Italy
Short description of the test	Detection of <i>Xylella fastidiosa</i> subsp. <i>pauca</i> ceppo CoDiRo by LAMP-PCR from DNA extracted from olive leaves and petioles
Date, reference of the validation report	2015-10-28 - Loreti S., Pucci N., Loconsole G., Modesti V, Lucchesi S.,Potere O., Saponari M 2017. Protocollo Diagnostico per XYLELLA FASTIDIOSA subsp. PAUCA ceppo CoDiRO. In Protocolli Diagnostici - ASPROPI- ISBN 9788899595722.pp. 241-278
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	<i>Xylella fastidiosa</i> subsp. <i>pauca</i> (XYLEFP)
Detection / identification	detection
Method(s)	Molecular Extraction DNA RNA Molecular LAMP Molecular LAMP (2)
Method: Molecular Extraction DNA RNA	
Reference of the test description	
Other information	
Other details on the test	DNA extraction by following Loconsole et al. (2014) (procedure B)
Method: Molecular LAMP	
Reference of the test description	
Kit	
Is a kit used	yes
Manufacturer name	ENBIOTECH

Specify the kit used	ENBIOTECH Xylella fastidiosa LAMP kit
Kit used following the manufacturer's instructions?	
Other information	
Method: Molecular LAMP (2)	
Reference of the test description	
Kit	
Is a kit used	yes
Manufacturer name	QUALIPLANTE
Specify the kit used	Isothermal PCR kit (ref bK.1/Xfas)
Kit used following the manufacturer's instructions?	
Other information	
Are the performance characteristics included in the EPPO diagnostic protocol?	yes
Performance Criteria :	
Organism 1.:	Xylella fastidiosa subsp. pauca(XYLEFP)
Analytical sensitivity	
What is smallest amount of target that can be detected reliably?	10 ² /10 ³ CFU/ml (Qualiplante SAS by using real-time PCR machine) 10 ³ CFU/ml (Enbiothec s.r.l. by using real-time PCR machine) 10 ¹⁻² CFU/ml (Enbiothec s.r.l. by using ic-gene)
Diagnostic sensitivity	
Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	90% LAMP-PCR (Enbiotech s.r.l.)
Standard test(s)	LAMP-PCR (Enbiotech s.r.l.)
Analytical specificity - exclusivity	
Number of non-target organisms tested	LAMP-PCR (Enbiotech s.r.l. kit) tested on the following bacterial strains: 3 Xanthomonas arboricola pv. pruni, 1 X. arboricola pv. juglandis, 2 X. arboricola pv. fragariae, 1 X. arboricola pv. corylina, 1 X. arboricola pv. celebensis), 1 X. campestris pv. campestris, 1 X. campestris pv. populi, 2 X. hortorum pv. pelargonii),3 Pseudomonas savastanoi pv. savastanoi, 1 P. marginalis, 4 P. syringae pv. syringae, 4 Brenneria (ssp. rubrifaciens, quercina, salicis, populi), 2 Pantoea stewartii, 1 Pantoea agglomerans, 1 Erwinia amylovora, 3 Agrobacterium tumefaciens, 2 Rhizobium vitis
Specificity value	No cross-reaction with LAMP-PCR (Enbiotech s.r.l. kit)
Diagnostic Specificity	
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	100% LAMP-PCR (Enbiotech s.r.l.)

Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	(Concordance) LAMP-PCR (Enbiotech s.r.l.): 85%
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	(Accordance) LAMP-PCR (Enbiotech s.r.l.): 91%
Test performance study	
Test performance study?	yes
Brief details of the test performance study and its output. It available, link to published article/report	<p>1. Two series of olive extracts spiked with ten fold dilution of Xylella fastidiosa CODiRo strain suspensions from 10^7 to 10^1 cfu/ml plus two healthy samples (16 samples in total) were tested in three different laboratories (CREA-PAV; CNR-IPSP; Plant Protection Service Lombardy) (NTC, healthy and infected olive extracts as control) for analytical sensitivity. 2. To check the diagnostic sensitivity and specificity, the accuracy, the repeatability and reproducibility, olive extract samples spiked with Xylella fastidiosa CODiRo strain suspensions at 10^6 cfu/ml (three repetitions), 10^4 cfu/ml (three repetitions), 10^3 cfu/ml (three repetitions), healthy olive extracts (three repetitions) for a total of 12 samples, were tested by the following TPS participants: 1. CREA-DC (N. Pucci; S. Loreti) 2. SELGE/CNR-IPSP/ DiSSPA-Uniba (M. Saponari, G. Loconsole; O. Potere) 3. PPS Piemonte (C. Morone, G. Mason) 4. PPS Friuli Venezia Giulia (G. Bianchi) 5. PPS Lombardia (F. Gaffuri) 6. PPS Emilia Romagna (A. Alessandrini; R. Gozzi) 7. PPS Trentino Alto Adige (V. Gualandri; L. Tessari) 8. PPS Marche (S. Nardi; S. Talevi) 9. PPS Liguria (M. Guelfi) 10. CIHEAM-IAMB (A.M. D'Onghia; M. Digiario) 11. CRSFA (F. Palmisano) 12. Centro di Sperimentazione Agraria e Forestale, Laimburg (A. Gallmetzer; A. Kraus) 13. Uni-MI (P. Casati) 14. Uni-CT (V. Catara) 15. PPS Toscana (D. Rizzo) 16. PPS Veneto (A. Saccardi; D. Pasqua di Bisceglie)</p>
Other information	
Any other information considered useful	<p>Accuracy: 92% (Enbiotech s.r.l.) Validation data were carried out by the Laboratories listed below, under the supervision of the reference laboratory CREA-PAV with the collaboration of CNR-IPSP: • CREA-PAV: Centro di Ricerca per la Patologia Vegetale CREA, Rome (Italy) • CNR-IPSP: Istituto per la Protezione Sostenibile delle Piante CNR, UOS Bari (Italy) • UNIBA: Dipartimento di Scienze del Suolo, della Pianta e degli Alimenti, Università degli Studi Aldo Moro, Bari (Italy); • Servizio Fitosanitario Regione Lombardia, Laboratorio Fitopatologico c/o Fondazione Minoprio 22070 Vertemate con Minoprio (CO) Italy</p>