

**EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION  
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES  
Summary sheet of validation data for a diagnostic test**

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

<b>Laboratory contact details</b>	Institute for Sustainable Plant Protection via Amendola, 122/D, 70126 Bari, Italy
<b>Short description of the test</b>	detection of <i>Xylella fastidiosa</i> in composite samples of herbaceous hosts
<b>Date, reference of the validation report</b>	2019-03-13 - 2019-03-13 Implementation of sampling procedures for testing composite samples for <i>Xylella fastidiosa</i> . POnTE - XF-ACTORS, 2nd Joint Annual Meeting: European Research on Emerging Plant Diseases. Valencia, 23-26 october 2018. Book of abstract: p. 63.
<b>Validation process according to EPPO Standard PM7/98?</b>	no
<b>Is the lab accredited for this test?</b>	yes
<b>Was the validated data generated in the framework of a project?</b>	
<b>Description of the test</b>	
<b>Organism(s)</b>	<i>Xylella fastidiosa</i> (XYLEFA)
<b>Detection / identification</b>	detection
<b>Method(s)</b>	Extraction Molecular Extraction DNA RNA Molecular Extraction DNA RNA (2) Molecular Extraction DNA RNA (3) Molecular real time PCR
<b>Method: Extraction</b>	
<b>Reference of the test description</b>	
<b>Other information</b>	
<b>Other details on the test</b>	The preparation of the samples of polygala is different from the description reported in the EPPO DP. W
<b>Method: Molecular Extraction DNA RNA</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>EPPO Diagnostic Protocol name</b>	PM 7/024 <i>Xylella fastidiosa</i> (version 4)
<b>Kit</b>	

<b>Is a kit used</b>	no
<b>Other information</b>	
<b>Other details on the test</b>	CTAB-based protocol
<b>Method: Molecular Extraction DNA RNA (2)</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>EPPO Diagnostic Protocol name</b>	PM 7/024 Xylella fastidiosa (version 4)
<b>Kit</b>	
<b>Is a kit used</b>	yes
<b>Manufacturer name</b>	QIAGEN
<b>Specify the kit used</b>	DNeasy mericon Food Kit
Kit used following the manufacturer's instructions?	no Modified DNeasy Mericon™ Food Standard Protocol
<b>Other information</b>	
<b>Method: Molecular Extraction DNA RNA (3)</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	no
<b>As or adapted from an IPPC diagnostic protocol</b>	no
<b>Kit</b>	
<b>Is a kit used</b>	yes
<b>Manufacturer name</b>	PROMEGA
<b>Specify the kit used</b>	Maxwell® RSC PureFood GMO and Authentication Kit
Kit used following the manufacturer's instructions?	
<b>Other information</b>	
<b>Method: Molecular real time PCR</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>EPPO Diagnostic Protocol name</b>	PM 7/024 Xylella fastidiosa (version 3)
<b>Name of the test</b>	Real-time PCR - simplex (Harper et al., 2010; erratum 2013)
<b>Is the test modified compared to the reference test</b>	yes BSA was not included in the amplification MIX
<b>Other information</b>	
<b>Other details on the test</b>	qPCR following the condition reported in Appendix 5 - Realtime PCR (Harper et al.,2010; erratum 2013) in PM 7/24 (3)

<b>Performance Criteria :</b>	
<b>Organism 1.:</b>	<b>Xylella fastidiosa(XYLEFA)</b>
<b>Analytical sensitivity</b>	
<b>What is smallest amount of target that can be detected reliably?</b>	One stem portion of 1,5-2 cm excised from an infected plant of periwinkle, in 40 gr of 1,5-2 cm stem portions of tomato plants
<b>Diagnostic sensitivity</b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	100% using all the 3 tests reported for the DNA extraction.
<b>Standard test(s)</b>	Standard tests reported in appendix 3 and 5 of PM 7/24 (3)
<b>Diagnostic Specificity</b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	100% using all the 3 tests reported for the DNA extraction.
<b>Specify the test(s)</b>	Standard tests reported in appendix 3 and 5 of PM 7/24 (3)
<b>Repeatability</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	100% using all the 3 tests reported for the DNA extraction.
<b>Test performance study</b>	
<b>Test performance study?</b>	no
<b>Other information</b>	
<b>Any other information considered useful</b>	These validation data were obtained by IPSP-CNR in collaboration with the Department of Soil, Plant and Food Science of the University of Bari (ITAY). For any additional detail, see the attached file.
The following complementary files are available online:	
	<ul style="list-style-type: none"> <li>• <a href="#">composite samples of herbaceous plants (tomato)</a></li> </ul>

Creation date: 2019-03-21 00:00:00 - Last update: 2022-01-25 14:17:14