EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity describes how validation should be conducted. It also includes definitions of performance criteria.

	•	
Laboratory contact details	Naktuinbouw Sotaweg 22, 2371 GD Roelofarendsveen, Netherlands	
Short description of the test	Detection of pospiviroids by realtime RT-PCR on tomato and pepper seeds, i.e. CEVd, CLVd, PCFVd, TASVd, TCDVd, TPMVd	
Date, reference of the validation report	2015-10-08 - TESTA Deliverable 5.3 Validated methods for viruses and viroids	
Validation process according to EPPO Standard PM7/98?	yes	
Is the lab accredited for this test?	yes	
Was the validated data generated in the framework of a project?	Other_project	
If yes, please specify	TESTA	
Description of the test		
Organism(s)	Pospiviroid (1POSPG)	
Detection / identification	detection	
Method(s)	Extraction Molecular Extraction DNA RNA Molecular real time RT PCR	
Method: Extraction		
Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
EPPO Diagnostic Protocol name	PM 7/138 Pospiviroids (genus Pospiviroid) (version 1)	
Name of the test	Seeds - Homogenization in GH+ buffer	
As or adapted from an IPPC diagnostic protocol	no	
Other information		
Other details on the test	Tomato seeds: crush seeds in buffer using stomacher Pepper seeds: crush seeds using genogrinder and subsequently add buffer	
Method: Molecular Extraction DNA RNA		

Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
EPPO Diagnostic Protocol name	PM 7/138 Pospiviroids (genus Pospiviroid) (version 1)	
As or adapted from an IPPC diagnostic protocol	no	
Kit		
Is a kit used	yes	
Manufacturer name	LGC	
Specify the kit used	sbeadex maxi plant	
Kit used following the manufacturer's instructions?		
Other information		
Other details on the test	RNA isolation using Kingfisher with Sbeadex kit	
Method: Molecular real time RT PCR		
Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
EPPO Diagnostic Protocol name	PM 7/138 Pospiviroids (genus Pospiviroid) (version 1)	
Name of the test	Real-time RT-PCR for seed testing (Naktuinbouw, 2017)	
As or adapted from an IPPC diagnostic protocol	no	
Other information		
Reaction type	Probe	
Other details on the test	Descriptions of the tests are available in complementary files available online (see links at the end of the sheet)	
Are the performance characteristics included in the EPPO diagnostic protocol?	yes	
Performance Criteria :		
Organism 1.:	Pospiviroid(1POSPG)	
Analytical sensitivity		
What is smallest amount of target that can be detected reliably?	For all seven viroids at least the 1000x dilution was detected and therefore the requirement detection of at least the 100x dilution was met. (Only for TPMVd not all 1000x dilutions were detected below the threshold of 32)	
Diagnostic sensitivity		
Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	100% for reaction mix A (PSTVd, PCFVd, TCDVd), note that primers and probes are similar to those used in the standard test. No data for reaction	

	mixes B, C and D. No data for pepper seeds.	
Standard test(s)	Detection of PSTVd and TCDVd in tomato seeds as described by Bakker et al (2015), EPPO Bulletin.	
Analytical specificity - inclusivity		
Number of strains/populations of target organisms tested	All 18 isolates of 7 species detected (TESTA report, Table 6).	
Specificity value	The analytical specificity was good since no false- negatives were observed for all primer sets and none of the non-target viroids and viruses reacted with the PCRs. Some acceptable cross-reactivity of TASVd isolates with the CEVd/CLVd primer mix (B) was observed. Objective of the seed assay is detect all relevant pospiviroids and identification of the pospiviroid is relatively less important.	
Analytical specificity - exclusivity		
Number of non-target organisms tested	No cross reactions with 29 isolates of other viruses and viroids tested (TESTA report, Table 7).	
Specificity value	Only cross-reactivity observed within pospiviroids, no cross-reactivity with other viroids or viruses	
<u>Diagnostic Specificity</u>		
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	100% for reaction mix A (PSTVd, PCFVd, TCDVd), note that primers and probes are similar to those used in the standard test. No data for reaction mixes B, C and D. No data for pepper seeds.	
Specify the test(s)	Detection of PSTVd and TCDVd in tomato seeds as described by Bakker et al (2015), EPPO Bulletin.	
Reproducibility		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for all target species	
Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for all target species	
Test performance study		
Test performance study?	no	
Other information		
Any other information considered useful	A proficiency test with four laboratories with naturally PSTVd-contaminated tomato seeds (1 PSTVd genotype) showed that the SPN-V043 2.0 method at Naktuinbouw did perform well. Multiple samples with only 10 PSTVd contaminated seeds amongst 990 healthy tomato seeds were detected.	
The following complementary files are available online:	 Test description Pepper seeds Test description tomato seeds Validation report Detection of pospiviroids by PCR in tomato seeds 	

Creation date: 2016-12-15 00:00:00 -	Last update: 2021-05-14 16:38:32