EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Dutch General Inspection Service (NAK) Randweg 14, 8304AS Emmeloord, Netherlands
Short description of the test	Detection of pospiviroid in potato leaves by real- time RT-PCR
Date, reference of the validation report	2011-11-01 - Rapport van de labvalidatie: aantonen van pospivoroïden in aardappelblad m.b.v. realtime RT-PCR
Validation process according to EPPO Standard PM7/98?	no
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	Pospiviroid (1POSPG)
Detection / identification	detection
Method(s)	Molecular real time RT PCR
Method: Molecular real time RT PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	no
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	M. Botermans, B.L.T.H. van de Vossenberg, J.Th.J. Verhoeven, M. Hooftman, R. Dekter, E.T.M. Meekes (2013). Development and validation of a real-time RT-PCR assay for generic detection of pospiviroids. Journal of Virological Methods: 187:43-50
Other information	
Are the performance characteristics included in the EPPO diagnostic protocol?	no
Performance Criteria :	
Organism 1.:	Pospiviroid(1POSPG)
Analytical sensitivity	

A) 1 infected leave in a pool of 100 healthy leaves B) sap of infected leave can be diluted 1: 2.500 in sap of healthy potato leaves	
Analytical specificity - inclusivity	
CEVd, CLVd, CSVd, MPVd, PCFVd, PSTVd, TASVd, TCDVd, TPMVd, IrVd	
Analytical specificity: highly specific for Pospiviroid species, no cross reactions with viruses commonly occurring in potato	
Analytical specificity - exclusivity	
PVY, PVYn, PVYoc, PVA, PVX, PVS, PVM, PVV, PLRV, PepMV, PMTV, APLV, APMoV, PBRSV.	
Reproducibility	
100%	
Repeatability	
100%	
Test performance study	
no	

Creation date: 2014-04-29 00:00:00 - Last update: 2021-05-21 15:41:39