

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Dutch General Inspection Service (NAK) Randweg 14, 8304AS Emmeloord, Netherlands
Short description of the test	Detection of pospiviroid in potato leaves by real-time RT-PCR
Date, reference of the validation report	2011-11-01 - Rapport van de labvalidatie: aantonen van pospivoroïden in aardappelblad m.b.v. real-time RT-PCR
Validation process according to EPPO Standard PM7/98?	no
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	Pospiviroid (1POSPG)
Detection / identification	detection
Method(s)	Molecular real time RT PCR
Method: Molecular real time RT PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	no
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	M. Botermans, B.L.T.H. van de Vossenbergh, J.Th.J. Verhoeven, M. Hooftman, R. Dekter, E.T.M. Meekes (2013). Development and validation of a real-time RT-PCR assay for generic detection of pospiviroids. Journal of Virological Methods: 187:43-50
Other information	
Are the performance characteristics included in the EPPO diagnostic protocol?	no
Performance Criteria :	
Organism 1.:	Pospiviroid(1POSPG)
Analytical sensitivity	

What is smallest amount of target that can be detected reliably?	A) 1 infected leave in a pool of 100 healthy leaves B) sap of infected leave can be diluted 1: 2.500 in sap of healthy potato leaves
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	CEVd, CLVd, CSVd, MPVd, PCFVd, PSTVd, TASVd, TCDVd, TPMVd, IrVd
Specificity value	Analytical specificity: highly specific for Pospiviroid species, no cross reactions with viruses commonly occurring in potato
Analytical specificity - exclusivity	
Number of non-target organisms tested	PVY, PVYn, PVYoc, PVA, PVX, PVS, PVM, PVV, PLRV, PepMV, PMTV, APLV, APMoV, PBRV.
Specificity value	
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%
Test performance study	
Test performance study?	no

Creation date: 2014-04-29 00:00:00 - Last update: 2021-05-21 15:41:39