

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France
Short description of the test	Detection of <i>Bursaphelenchus xylophilus</i> in wood extract with real-time PCR Leal et al. 2005
Date, reference of the validation report	2011-02-01 - Anses 2011 Rapport d'évaluation d'outils moléculaires de détection de <i>Bursaphelenchus xylophilus</i> sur extrait de bois
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	<i>Bursaphelenchus xylophilus</i> (BURSXY)
Detection / identification	detection
Method(s)	Molecular real time PCR
Method: Molecular real time PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	no
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	Leal I, Green M, Allen E, Humble L, Rott M (2005) An effective PCR-based diagnostic method for the detection of <i>Bursaphelenchus xylophilus</i> in wood samples from lodgepole pine. <i>Nematology</i> Vol.7(6), 833-842
Other information	
Reaction type	Probe
Performance Criteria :	
Organism 1.:	<i>Bursaphelenchus xylophilus</i>(BURSXY)
Analytical sensitivity	

What is smallest amount of target that can be detected reliably?	one nematode
<u>Diagnostic sensitivity</u>	
Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	1
Standard test(s)	morphology
<u>Analytical specificity - inclusivity</u>	
Number of strains/populations of target organisms tested	7 populations (see table1)
Specificity value	100%
<u>Analytical specificity - exclusivity</u>	
Number of non-target organisms tested	15 populations (see table1)
Specificity value	100% - no cross reactions
<u>Diagnostic Specificity</u>	
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	99% (6 false positives results/490 uninfested samples and 1 false negative result /13 infested samples)
Specify the test(s)	morphology
<u>Reproducibility</u>	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for one single nematode
<u>Repeatability</u>	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for one single nematode
<u>Test performance study</u>	
Test performance study?	no
<u>Other information</u>	
Any other information considered useful	The full report is available upon request to the laboratory. French version only
The following complementary files are available online:	<ul style="list-style-type: none"> • Table 1 List of species and population tested • Table 2 comparison of different PCR tests B xylophilus identification

Creation date: 2012-09-03 00:00:00 - Last update: 2021-06-18 10:34:11