

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France
Short description of the test	Identification of <i>Bursaphelenchus xylophilus</i> by nested PCR (Takeuchi et al. 2005)
Date, reference of the validation report	2011-05-01 - Validation report may 2011
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	<i>Bursaphelenchus xylophilus</i> (BURSXY)
Detection / identification	identification
Matrix(ces) tested	Specimen Isolated nematodes
Method(s)	Molecular Conventional PCR
Method: Molecular Conventional PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	no
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	Takeuchi Y, Kansaki N & Futai K (2005) A nested PCR-based method for detecting the pinewood nematode, <i>Bursaphelenchus xylophilus</i> , from pinewood. <i>Nematology</i> 7(5), 775-782. Not included in appendix of PM7/4 (2).
Other information	
Reaction type	Nested
Performance Criteria :	
Organism 1.:	<i>Bursaphelenchus xylophilus</i>(BURSXY)
Analytical sensitivity	
What is the smallest amount of target that	5 nematodes

can be detected reliably?	
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	7 populations
Specificity value	100%
Analytical specificity - exclusivity	
Number of non-target organisms tested	15 populations
Specificity value	cross reactions with some populations of <i>B. mucronatus</i> and <i>B. sp</i> from France and China
Cross-reacts with	<i>Bursaphelenchus sp.</i> <i>Bursaphelenchus mucronatus</i>
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for 5 individuals
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for 5 individuals
Test performance study	
Test performance study?	no
Other information	
Any other information considered useful	The full report is available upon request to the laboratory. French version only

Creation date: 2012-12-26 00:00:00 - Last update: 2021-06-15 16:13:00