

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France
Short description of the test	Detection of Bursaphelenchus xylophilus in wood extract with PCR RFLP Burgersmeister et al. 2005
Date, reference of the validation report	2011-05-01 - Evaluation d'outils moléculaires d'identification de Bursaphelenchus xylophilus sur individus isolés
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	no
Was the validated data generated in the framework of a project?	no
Description of the test	
Organism(s)	Bursaphelenchus xylophilus (BURSXY)
Detection / identification	detection
Method(s)	Molecular PCR-RFLP
Method: Molecular PCR-RFLP	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	yes
EPPO Diagnostic Protocol name	PM 7/004 Bursaphelenchus xylophilus (version 3)
Name of the test	ITS RFLP PCR (Burgermeister et al., 2009)
As or adapted from an IPPC diagnostic protocol	no
Other information	
Performance Criteria :	
Organism 1.:	Bursaphelenchus xylophilus(BURSXY)
Analytical sensitivity	
What is smallest amount of target that can be detected reliably?	5 nematodes
Diagnostic sensitivity	

Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	not performed
Standard test(s)	not relevant
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	7 populations (see table1)
Specificity value	100%
Analytical specificity - exclusivity	
Number of non-target organisms tested	15 populations (see table1)
Specificity value	100% - no cross reaction
Diagnostic Specificity	
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	not performed
Specify the test(s)	not relevant
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for 5 individuals
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%
Test performance study	
Test performance study?	no
Other information	
Any other information considered useful	See table 2. The full report is available upon request to the laboratory. French version only
The following complementary files are available online:	<ul style="list-style-type: none"> • Table 1 Specificity_PCR

Creation date: 2012-09-04 00:00:00 - Last update: 2021-06-15 16:12:19