

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

| | |
|---|--|
| Laboratory contact details | Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France |
| Short description of the test | Detection of <i>Bursaphelenchus xylophilus</i> in wood extract with PCR RFLP Burgersmeister et al. 2005 |
| Date, reference of the validation report | 2011-05-01 - Evaluation d'outils moléculaires d'identification de <i>Bursaphelenchus xylophilus</i> sur individus isolés |
| Validation process according to EPPO Standard PM7/98? | yes |
| Is the lab accredited for this test? | no |
| Was the validated data generated in the framework of a project? | no |
| Description of the test | |
| Organism(s) | <i>Bursaphelenchus xylophilus</i> (BURSXY) |
| Detection / identification | detection |
| Matrix(ces) tested | Specimen Nematodes |
| Method(s) | Molecular PCR-RFLP |
| Method: Molecular PCR-RFLP | |
| Reference of the test description | |
| As or adapted from an EPPO diagnostic protocol | yes |
| EPPO Diagnostic Protocol name | PM 7/004 <i>Bursaphelenchus xylophilus</i> (version 3) |
| Name of the test | ITS RFLP PCR (Burgermeister et al., 2009) |
| As or adapted from an IPPC diagnostic protocol | no |
| Other information | |
| Performance Criteria : | |
| Organism 1.: | <i>Bursaphelenchus xylophilus</i>(BURSXY) |
| Analytical sensitivity | |
| What is the smallest amount of target that can be detected reliably? | 5 nematodes |

| | |
|--|---|
| Diagnostic sensitivity | |
| Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98 | not performed |
| Standard test(s) | not relevant |
| Analytical specificity - inclusivity | |
| Number of strains/populations of target organisms tested | 7 populations (see table1) |
| Specificity value | 100% |
| Analytical specificity - exclusivity | |
| Number of non-target organisms tested | 15 populations (see table1) |
| Specificity value | 100% - no cross reaction |
| Diagnostic Specificity | |
| Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test | not performed |
| Specify the test(s) | not relevant |
| Reproducibility | |
| Provide the calculated % of agreement for a given level of the pest (see PM 7/98) | 100% for 5 individuals |
| Repeatability | |
| Provide the calculated % of agreement for a given level of the pest (see PM 7/98) | 100% |
| Test performance study | |
| Test performance study? | no |
| Other information | |
| Any other information considered useful | See table 2. The full report is available upon request to the laboratory. French version only |
| | |
| The following complementary files are available online: | <ul style="list-style-type: none"> • Table 1 Specificity_PCR |

Creation date: 2012-09-04 00:00:00 - Last update: 2021-06-15 16:12:19