

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France
Short description of the test	Specied specific PCR detection for <i>Meloidogyne chitwoodi</i> , <i>M. fallax</i> and <i>M. hapla</i>
Date, reference of the validation report	2010-09-01 - Validation report - Septembre 2010
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	yes
Was the validated data generated in the framework of a project?	
Description of the test	
Organism(s)	<i>Meloidogyne chitwoodi</i> (MELGCH) <i>Meloidogyne fallax</i> (MELGFA)
Detection / identification	detection
Matrix(ces) tested	Specimen isolated nematodes
Method(s)	Molecular Conventional PCR
Method: Molecular Conventional PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	yes
EPPO Diagnostic Protocol name	PM 7/041 <i>Meloidogyne chitwoodi</i> and <i>M. fallax</i> (version 2)
Name of the test	PCR Wishart et al. 2002
Is the test modified compared to the reference test	yes dNTPs concentration reduced
Other information	
Other details on the test	Species specific PCR (IGS region)
Are the performance characteristics included in the EPPO diagnostic protocol?	yes
Performance Criteria :	
Organism 1.:	<i>Meloidogyne chitwoodi</i>(MELGCH)

Analytical sensitivity	
What is the smallest amount of target that can be detected reliably?	1 J2 for M. chitwoodi
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	5 populations for M. chitwoodi (see annex 1 of validation report for details)
Specificity value	
Analytical specificity - exclusivity	
Number of non-target organisms tested	28 populations of nematodes (see Annex 1 of validation report for details)
Specificity value	Detected as M. chitwoodi for some of the replicates : M. javanica (1 population), M. enterolobii (1 population), Heterodera schachtii and Xiphinema sp.
Cross-reacts with	Meloidogyne javanica Meloidogyne enterolobii Heterodera schachtii Xiphinema sp.
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for M. chitwoodi
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100 % for M. chitwoodi
Organism 2.:	Meloidogyne fallax(MELGFA)
Analytical sensitivity	
What is the smallest amount of target that can be detected reliably?	1 J2 for M. fallax
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	2 populations for M. fallax (see annex 1 of validation report for details)
Specificity value	100% for M. fallax
Analytical specificity - exclusivity	
Number of non-target organisms tested	28 populations of nematodes (see Annex 1 of validation report for details)
Specificity value	100% for M. fallax
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for M. fallax
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100 % for M. fallax

Test performance study	
Test performance study?	no
Other information	
Any other information considered useful	The full report is available upon request to the laboratory.

Creation date: 2012-12-26 00:00:00 - Last update: 2021-05-20 17:45:48