## EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Nematology Unit Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France	
Short description of the test	Specied specific PCR detection for Meloidogyne chitwoodi, M. fallax and M. hapla	
Date, reference of the validation report	2010-09-01 - Validation report - Septembre 2010	
Validation process according to EPPO Standard PM7/98?	yes	
Is the lab accredited for this test?	yes	
Was the validated data generated in the framework of a project?		
Description of the test		
Organism(s)	Meloidogyne chitwoodi (MELGCH) Meloidogyne fallax (MELGFA)	
Detection / identification	detection	
Method(s)	Molecular Conventional PCR	
Method: Molecular Conventional PCR		
Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
EPPO Diagnostic Protocol name	PM 7/041 Meloidogyne chitwoodi and M. fallax (version 2)	
Name of the test	PCR Wishart et al. 2002	
Is the test modified compared to the reference test	yes dNTPs cencentration reduced	
Other information		
Other details on the test	Species specific PCR (IGS region)	
Are the performance characteristics included in the EPPO diagnostic protocol?	yes	
Performance Criteria :		
Organism 1.:	Meloidogyne chitwoodi(MELGCH)	
Analytical sensitivity		

What is smallest amount of target that can be detected reliably?	1 J2 for M. chitwoodi	
Analytical specificity - inclusivity		
Number of strains/populations of target organisms tested	5 populations for M. chotwoodi (see annex 1 of validation report for details)	
Specificity value		
Analytical specificity - exclusivity		
Number of non-target organisms tested	28 populations of nematodes (see Annex 1 of validation report for details)	
Specificity value	Detected as M. chitwoodi for some of the replicates : M. javanica (1 population), M. enterolobii (1 population), Heterodera schachtii and Xiphinema sp.	
Cross reacts with	Meloidogyne javanica Meloidogyne enterolobii Heterodera schachtii Xiphinema sp.	
Reproducibility		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for M. chitwoodi	
Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100 % for M. chitwoodi	
Organism 2.:	Meloidogyne fallax(MELGFA)	
Analytical sensitivity		
What is smallest amount of target that can be detected reliably?	1 J2 for M. fallax	
Analytical specificity - inclusivity		
Number of strains/populations of target organisms tested	2 populations for M. falalx (see annex 1 of validation report for details)	
Specificity value	100% for M. fallax	
Analytical specificity - exclusivity		
Number of non-target organisms tested	28 populations of nematodes (see Annex 1 of validation report for details)	
Specificity value	100% for M. fallax	
Reproducibility		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for M. fallax	
Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100 % for M. fallax	

Test performance study?	no	
Other information		
Any other information considered useful	The full report is available upon request to the laboratory.	

Creation date: 2012-12-26 00:00:00 - Last update: 2021-05-20 17:45:48