

**EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION**  
**ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES**  
**Summary sheet of validation data for a diagnostic test**

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

<b>Laboratory contact details</b>	Anses Plant Health Laboratory - Pests and Tropical Pathogens Unit Pôle de Protection des Plantes, 7 Chemin de l'IRAT, 97410 Saint Pierre, France
<b>Short description of the test</b>	Detection and identification of 'Candidatus Liberibacter asiaticus' and 'Candidatus Liberibacter africanus' by Molecular real time PCR (according to Morgan et al., 2012 using SYBR) in Citrus sp. leaves
<b>Date, reference of the validation report</b>	2020-07-10 - HLB_qPCR_EUPHRESCO-2016-A-232
<b>Link to other validation data</b>	- HLB_qPCR_EUPHRESCO-2016-A-232 Detection and identification of 'Candidatus Liberibacter asiaticus' and 'Candidatus Liberibacter africanus' by Molecular real time PCR (according to Morgan et al., 2012 using SYBR) in Citrus sp. leaves
<b>Validation process according to EPPO Standard PM7/98?</b>	yes
<b>Is the lab accredited for this test?</b>	no
<b>Was the validated data generated in the framework of a project?</b>	Euphresco
<b>If yes, please specify</b>	2016-A-232
<b>Description of the test</b>	
<b>Organism(s)</b>	'Candidatus Liberibacter asiaticus'(LIBEAS) 'Candidatus Liberibacter africanus'(LIBEAF)
<b>Detection / identification</b>	detection and identification
<b>Method(s)</b>	Molecular Extraction DNA RNA Molecular real time PCR
<b>Method: Molecular Extraction DNA RNA</b>	
<b>Reference of the test description</b>	
<b>Kit</b>	
<b>Is a kit used</b>	yes
<b>Manufacturer name</b>	QIAGEN
<b>Specify the kit used</b>	DNeasy Plant Mini Kit
<b>Kit used following the manufacturer's instructions?</b>	yes DNA extraction was performed on ground citrus leaves using the DNeasy Plant Mini Kit (Qiagen, Germantown, MD, USA) following the

	manufacturer's recommendations.
<b>Other information</b>	
<b>Other details on the test</b>	Ground using a HOMEX 6 homogenizer (Bioreba AG, Reinach, Switzerland) with 5 mL of extraction buffer (pH = 8): 50 mM Sigma 7-9® TRIS (Merck KGaA, Darmstadt, Germany); 5 mM EDTA (Merck KGaA); and 1% sodium dodecyl sulfate (Merck KGaA).
<b>Method: Molecular real time PCR</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>EPPO Diagnostic Protocol name</b>	PM 7/121 ' <i>Candidatus</i> Liberibacter africanus', ' <i>Candidatus</i> Liberibacter americanus' and ' <i>Candidatus</i> Liberibacter asiaticus' (version 1)
<b>Name of the test</b>	Real-time PCR targeting hyvI/hyvII gene (according to Morgan et al. 2012)
<b>As or adapted from an IPPC diagnostic protocol</b>	no
<b>Is the test modified compared to the reference test</b>	yes Use of SYBR green instead of Taqman probe
<b>Kit</b>	
<b>Is a kit used</b>	no
<b>Other information</b>	
<b>Reaction type</b>	Simplex
<b>Are the performance characteristics included in the EPPO diagnostic protocol?</b>	no
<b>Performance Criteria :</b>	
<b>Organism 1.:</b>	' <i>Candidatus Liberibacter asiaticus</i> '(LIBEAS)
<b>Diagnostic sensitivity</b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	0.626 (DSE=PA/N+, Fav. Hypothesis, considering CLas and CLaf samples)
<b>Diagnostic Specificity</b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	0.980 (DSP=NA/N-, Fav. Hypothesis, considering CLas and CLaf samples)
<b>Reproducibility</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	CO=0.952 (considering CLas and CLaf samples)
<b>Repeatability</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	DA=0.976 (considering CLas and CLaf samples)

<b>Organism 2.:</b>	<b>'Candidatus Liberibacter africanus'(LIBEAF)</b>
<b><u>Diagnostic sensitivity</u></b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	0.626 (DSE=PA/N+, Fav. Hypothesis, considering CLas and CLaf samples)
<b>Cross reacts with</b>	'Candidatus Liberibacter solanacearum'
<b><u>Diagnostic Specificity</u></b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	0.980 (DSP=NA/N-, Fav. Hypothesis, considering CLas and CLaf samples)
<b><u>Reproducibility</u></b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	CO=0.952 (considering CLas and CLaf samples)
<b><u>Repeatability</u></b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	DA=0.976 (considering CLas and CLaf samples)
<b><u>Test performance study</u></b>	
<b>Test performance study?</b>	yes
<b>Brief details of the test performance study and its output.It available, link to published article/report</b>	Test performance study organized in the framework of a EUPHRESKO project involving 8 international laboratories.
<b><u>Other information</u></b>	
<b>Any other information considered useful</b>	Publication available at: <a href="https://link.springer.com/content/pdf/10.1007/s10658-020-02052-3.pdf">https://link.springer.com/content/pdf/10.1007/s10658-020-02052-3.pdf</a> Cellier, G., C. Redondo, J. Cubero, M. Roselló, E. de Andrade, L. Cruz, E. Ince, H. N. Yildiz, P. G. Güler, A. M. D'Onghia, T. Yaseen, K. Djelouah, E. Metz-Verschure, F. Gaffuri, R. A. Gottsberger, and B. Giovani. 2020. "Comparison of the performance of the main real-time and conventional PCR detection tests for 'Candidatus Liberibacter' spp., plant pathogenic bacteria causing the Huanglongbing disease in Citrus spp." European Journal of Plant Pathology. doi: 10.1007/s10658-020-02052-3.

Creation date: 2020-08-18 15:13:09 - Last update: 2020-09-14 15:08:47