

**EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION**  
**ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES**  
**Summary sheet of validation data for a diagnostic test**

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

<b>Laboratory contact details</b>	National Institute of Biology, Department of Biotechnology and Systems Biology Vecna pot 111, 1000 Ljubljana, Slovenia
<b>Short description of the test</b>	Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Boonham et al. 2002) in symptomatic tomato leaves (TPS)
<b>Date, reference of the validation report</b>	2020-12-14 - TSWV V1.0
<b>Link to other validation data</b>	<ul style="list-style-type: none"> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Mortimer-Jones et al. 2009) in symptomatic tomato leaves (prevalidation study)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Boonham et al. 2002) in symptomatic tomato leaves (prevalidation study)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Roberts et al. 2000) in symptomatic tomato leaves (prevalidation study)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by RT-PCR in symptomatic tomato leaves (prevalidation study)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by RT-PCR in symptomatic tomato leaves (TPS)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Mortimer-Jones et al. 2009) in symptomatic tomato leaves (TPS)</li> <li>- TSWV V1.0 Detection and identification of tomato spotted wilt tospovirus by real-time RT-qPCR (test adapted from Roberts et al. 2000) in symptomatic tomato leaves (TPS)</li> </ul>
<b>Validation process according to EPPO Standard PM7/98?</b>	yes
<b>Is the lab accredited for this test?</b>	no
<b>Was the validated data generated in the framework of a project?</b>	Other_project
<b>If yes, please specify</b>	VALITEST
<b>Description of the test</b>	
<b>Organism(s)</b>	Tomato spotted wilt virus(TSWV00)

<b>Detection / identification</b>	detection and identification
<b>Matrix(ces) tested</b>	Leaves different tospovirus isolates from DSMZ collection spiked into tomato leaf extract
<b>Plant species tested</b>	Solanum lycopersicum
<b>Method(s)</b>	Molecular Extraction DNA RNA Molecular real time RT PCR
<b>Method: Molecular Extraction DNA RNA</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>New test being considered for inclusion in the next version of the EPPO diagnostic protocol?</b>	
<b>EPPO Diagnostic Protocol name</b>	PM 7/139 Tospoviruses (genus Orthotospovirus) (version 1)
<b>Name of the test</b>	
<b>As or adapted from an IPPC diagnostic protocol</b>	no
<b>Is the test modified compared to the reference test</b>	no
<b>Kit</b>	
<b>Is a kit used</b>	yes
<b>Manufacturer name</b>	QIAGEN
<b>Specify the kit used</b>	RNeasy Plant Mini Kit
Kit used following the manufacturer's instructions?	no Extraction was performed as described in Appendix 3 of PM 7/139 (1). Samples were extracted by placing the 0.6 g of freeze dried leaves in Bioreba extraction bag and adding 3 ml of PBS-extraction buffer. An aliquot of 100 µL was send to TPS participants. TPS participants added 450 µL of RLT buffer (without β-mercaptoethanol) (Qiagen) and continued with the manufacturer's procedure. Total RNA was eluted twice with 50 µL (total of 100 µL) of RNase-free water pre-warmed to 65°C. Undiluted RNA was used for testing.
<b>Other information</b>	
<b>Other details on the test</b>	
<b>Method: Molecular real time RT PCR</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	yes
<b>New test being considered for inclusion in the next version of the EPPO diagnostic protocol?</b>	
<b>EPPO Diagnostic Protocol name</b>	PM 7/139 Tospoviruses (genus Orthotospovirus) (version 1)

<b>Name of the test</b>	Real-time RT-PCR for TSWV (Boonham et al., 2002)
<b>As or adapted from an IPPC diagnostic protocol</b>	no
<b>Is the test modified compared to the reference test</b>	no
<b>Kit</b>	
<b>Is a kit used</b>	yes
<b>Manufacturer name</b>	Ambion (Thermo Scientific)
<b>Specify the kit used</b>	AgPath-IDTM One-step RT-PCR kit
Kit used following the manufacturer's instructions?	no Final reaction volume was 10 µL. Final concentration of primers was 0.9 µM and final concentration of probe was 0.2 µM. Primer concentration and cycling conditions were the same for all used kits. Cycling conditions were: 48°C for 10 min, 95 °C for 10 min, then 45 cycles at 95 °C for 15 s and 60 °C for 1 min.
<b>Other information</b>	
<b>Reaction type</b>	Simplex
<b>Other details on the test</b>	kit from Ambion (Thermo Scientific) was used by 10 laboratories; kit from VWR, Quantabio (Quanta OneStep Toughmix qScript™ XLT One-StepRT-qPCR ToughMix®   VWR) was used by one laboratory; kit from Invitrogen/Thermo Fisher Scientific (PCR SuperScript™ III One-Step RT-PCR System with Platinum™ Taq DNA Polymerase) was used by one laboratory; kit from Quantabio (Ultraplex 1-step Toughmix (4X)) was used by one laboratory. No impact was observed when using different kits.
<b>Are the performance characteristics included in the EPPO diagnostic protocol?</b>	
<b>Performance Criteria :</b>	
<b>Organism 1.:</b>	<b>Tomato spotted wilt virus(TSWV00)</b>
<b>Analytical sensitivity</b>	
<b>What is smallest amount of target that can be detected reliably?</b>	at least to 1,000,000x dilution of isolate TSWV-PV-1175 (level of agreement between experiments: 100%)
<b>Diagnostic sensitivity</b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	99.50%
<b>Standard test(s)</b>	Calculation was done on the basis of expected results which was for molecular test the same as health status of the samples (1,000,000x and 100,000x diluted isolate TSWV-PV-1175 was expected to be positive). Additionally, in the TPS study, 1,000x diluted TSWV-PV-0182 isolate and 100,000x diluted TSWV-PV-0389 isolate have been used as medium and as low concentrated TSWV

	positive sample, respectively.
<b>Analytical specificity - inclusivity</b>	
<b>Number of strains/populations of target organisms tested</b>	3 TSWV isolates from DSMZ collection (PV-1175, PV-0182, PV-0389)
<b>Specificity value</b>	/
<b>Analytical specificity - exclusivity</b>	
<b>Number of non-target organisms tested</b>	3 healthy tomato samples, 5 other tospovirus species (ANSV00 isolate PV-1027; CSNV00 isolate PV-0529; GRSV00 isolate PV-0205; INSV00 isolate PV-0281; TCSV00 isolate PV-0390)
<b>Specificity value</b>	/
<b>Cross reacts with</b>	
<b>Diagnostic Specificity</b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	96.20%
<b>Specify the test(s)</b>	Calculation was done based on the results of samples listed in sections inclusivity and exclusivity. One TSWV isolate was tested as several dilutions (see section diagnostic sensitivity) and also these results are included in calculation of diagnostic specificity.
<b>Reproducibility</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	98.30%
<b>Repeatability</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	96.60%
<b>Test performance study</b>	
<b>Test performance study?</b>	yes
<b>Brief details of the test performance study and its output. It available, link to published article/report</b>	Preparation for test performance study organized in the framework of the VALITEST project
<b>Other information</b>	
<b>Any other information considered useful</b>	Test performance study organized in the framework of the VALITEST project involving 15 laboratories from 12 countries. Results of two laboratories have been excluded from calculation of diagnostic parameters due to non-conforming results of healthy tomato samples. Full validation report is available on request (send the e-mail to: niblabfito@nib.si and natasa.mehle@nib.si).

Creation date: 2020-12-31 21:42:56 - Last update: 2021-09-10 15:57:10