## EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Anses Plant Health Laboratory - Bacteriology, Virology and GMO Unit 7 rue Jean Dixméras, 49044 Angers, France	
Short description of the test	Detection of Xylella fastidiosa by molecular real time PCR in vectors	
Date, reference of the validation report	2024-10-23 - Détection de Xylella fastidiosa par PCR en temps réel sur insectes vecteurs MA065ver02	
Validation process according to EPPO Standard PM7/98?	yes	
Is the lab accredited for this test?	yes	
Was the validated data generated in the framework of a project?	no	
Description of the test		
Organism(s)	Xylella fastidiosa (XYLEFA)	
Detection / identification	detection	
Method(s)	Molecular Extraction DNA RNA Molecular real time PCR	
Method: Molecular Extraction DNA RNA		
Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
New test being considered for inclusion in the next version of the EPPO diagnostic protocol?	no	
EPPO Diagnostic Protocol name	PM 7/024 Xylella fastidiosa (version 5)	
As or adapted from an IPPC diagnostic protocol	no	
Is the test modified compared to the reference test	no	
Kit		
Is a kit used	yes	
Manufacturer name	BIONOBILE	
Specify the kit used	QuickPick™ SML Plant DNA	

Kit used following the manufacturer's instructions?	yes	
Other information		
Method: Molecular real time PCR		
Reference of the test description		
As or adapted from an EPPO diagnostic protocol	yes	
New test being considered for inclusion in the next version of the EPPO diagnostic protocol?	no	
EPPO Diagnostic Protocol name	PM 7/024 Xylella fastidiosa (version 5)	
Name of the test	Real-time PCR (Harper et al., 2010; erratum 2013)	
As or adapted from an IPPC diagnostic protocol	no	
Is the test modified compared to the reference test	no	
Kit		
Is a kit used	yes	
Manufacturer name	Applied Biosystems	
Specify the kit used	TaqMan™ Fast Universal Master Mix (2X), no AmpErase™ UNG	
Kit used following the manufacturer's instructions?	yes	
Other information		
Reaction type	Duplex	
Other details on the test	Duplex real time Harper et al., 2010 / loos et al., 2009 (internal control RNAr 18S)	
Performance Criteria :		
Organism 1.:	Xylella fastidiosa(XYLEFA)	
Analytical sensitivity		
What is smallest amount of target that can be detected reliably?	2.10^3 cell./head with a detection rate of 100%	
<u>Diagnostic sensitivity</u>		
Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98	100%	
Analytical specificity - inclusivity		
Number of strains/populations of target organisms tested	100%	
Specificity value		
<u>Diagnostic Specificity</u>		
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test	100%	

Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100%	
Test performance study		
Test performance study?	no	
The following complementary files are available online:	Validation report	

Creation date: 2025-03-17 16:13:22 - Last update: 2025-03-17 16:24:41