

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Laboratory contact details	Netherlands Institute for Vectors, Invasive plants and Plant health P.O. Box 9102, 6700 HC Wageningen, Netherlands
Short description of the test	This validation data is for generic detection of phytoplasmas. Phytoplasmas can be detected using real time PCR. For identification see validation sheet 555.
Date, reference of the validation report	2022-07-21 - 2020.molbio.012
Link to other validation data	- 2020.molbio.012 This validation data is for generic detection and identification of phytoplasmas. Phytoplasmas can be detected using conventional nested PCR. The conventional (nested) PCR product is purified and finally sequenced using HTS. For identification see validation sheet 571.
Validation process according to EPPO Standard PM7/98?	yes
Is the lab accredited for this test?	yes
Was the validated data generated in the framework of a project?	no
Description of the test	
Organism(s)	'Candidatus Phytoplasma aurantifolia' (PHYPAF) 'Candidatus Phytoplasma brasiliense' (PHYPBR) 'Candidatus Phytoplasma fraxini' (PHYPPFR) 'Candidatus Phytoplasma oryzae' (PHYPOR) 'Candidatus Phytoplasma phoenicium' (PHYPPH) 'Candidatus Phytoplasma pruni' (PHYPPN) 'Candidatus Phytoplasma pyri' (PHYPPY) 'Candidatus Phytoplasma solani' (PHYPSO) 'Candidatus Phytoplasma trifolii' (PHYPTR) 'Candidatus Phytoplasma ulmi' (PHYPUL) 'Candidatus Phytoplasma ziziphi' (PHYPZI) Grapevine flavescence dorée phytoplasma (PHYP64) 'Candidatus Phytoplasma americanum' (PHYPAE) 'Candidatus Phytoplasma palmicola' (PHYPPPL) 'Candidatus Phytoplasma palmae' (PHYPPA) Phytoplasma (1PHYPG) 'Candidatus Phytoplasma fragariae' (PHYPPFG)
Detection / identification	detection
Matrix(ces) tested	Leaves, Other Tests validated on: - Vitis sp. leaves

	(veins) provided by NIB - Catharanthus DNA provided by University of Bologna - Solanum lycopersicum DNA provided by University of Bologna - Cocos nucifera DNA provided by University of Bologna
Method(s)	Molecular Extraction DNA RNA Molecular real time PCR
Method: Molecular Extraction DNA RNA	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	yes
EPPO Diagnostic Protocol name	PM 7/133 Generic detection of phytoplasmas (version 1)
As or adapted from an IPPC diagnostic protocol	no
Is the test modified compared to the reference test	no
Kit	
Is a kit used	yes
Manufacturer name	QIAGEN
Specify the kit used	DNeasy Plant Mini Kit
Kit used following the manufacturer's instructions?	yes
Other information	
Method: Molecular real time PCR	
Reference of the test description	
As or adapted from an EPPO diagnostic protocol	yes
EPPO Diagnostic Protocol name	PM 7/133 Generic detection of phytoplasmas (version 1)
Name of the test	Real-time PCR for the generic detection of phytoplasmas (Hodgetts et al., 2009)
As or adapted from an IPPC diagnostic protocol	yes
IPPC diagnostic Protocol name	ISPM 27 Annex 12 DP 12: Phytoplasmas (version 2016)
Name of the test	Real-time PCR
Is the test modified compared to the reference test	no
Kit	
Is a kit used	no
Other information	
Reaction type	Simplex

Other details on the test	Appendix 4 real-time PCR for the generic detection of phytoplasmas (Hodgetts et al. 2009)
Performance Criteria :	
Organism 1.:	'Candidatus Phytoplasma aurantifolia'(PHYPAF)
Organism 2.:	'Candidatus Phytoplasma brasiliense'(PHYPPBR)
Organism 3.:	'Candidatus Phytoplasma fraxini'(PHYPPFR)
Organism 4.:	'Candidatus Phytoplasma oryzae'(PHYPPOR)
Organism 5.:	'Candidatus Phytoplasma phoenicium'(PHYPPH)
Organism 6.:	'Candidatus Phytoplasma pruni'(PHYPPN)
Organism 7.:	'Candidatus Phytoplasma pyri'(PHYPPY)
Organism 8.:	'Candidatus Phytoplasma solani'(PHYPPSO)
Organism 9.:	'Candidatus Phytoplasma trifolii'(PHYPPTR)
Organism 10.:	'Candidatus Phytoplasma ulmi'(PHYPPUL)
Organism 11.:	'Candidatus Phytoplasma ziziphi'(PHYPPZI)
Organism 12.:	Grapevine flavescence dorée phytoplasma(PHYPP64)
Organism 13.:	'Candidatus Phytoplasma americanum'(PHYPPAE)
Organism 14.:	'Candidatus Phytoplasma palmicola'(PHYPPPL)
Organism 15.:	'Candidatus Phytoplasma palmae'(PHYPPPA)
Organism 16.:	Phytoplasma(1PHYPPG)
Analytical sensitivity	
What is the smallest amount of target that can be detected reliably?	Grapevine flavescence doree phytoplasma is detected at 10 ³ dilution in the real time PCR (Hodgetts et al. 2009). GFDP is detected at 10 ² dilution in the conventional nested PCR (Lee et al., 1993 and Gundersen & Lee, 1996).
Analytical specificity - inclusivity	
Number of strains/populations of target organisms tested	see annex 4 (validation report) and table 1 in (addendum). Ca. Phytoplasma aurantifolia Ca. Phytoplasma aurantifolia-related strain SPLL Ca. Phytoplasma brasiliense Ca. Phytoplasma fraxini-reference strain Ca. Phytoplasma oryzae Ca. Phytoplasma phoenicium Ca. Phytoplasma pruni Ca. Phytoplasma pyri Ca. Phytoplasma solani Ca. Phytoplasma trifolii Ca. Phytoplasma ulmi Ca. Phytoplasma ulmi Ca. Phytoplasma ziziphi Grapevine flavescence doree phytoplasma Grapevine flavescence doree phytoplasma GFDP Map-FD1 Grapevine flavescence doree phytoplasma GFDP Map-FD2 Grapevine flavescence doree phytoplasma GFDP Map-FD2 PEY05 Grapevine flavescence doree phytoplasma GFDP

	Map-FD3 Ca. Phytoplasma americanum Ca. Phytoplasma palmae Ca. Phytoplasma palmicola
Specificity value	100
Reproducibility	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100
Repeatability	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100
Organism 17.:	'Candidatus Phytoplasma fragariae'(PHYPFG)
Test performance study	
Test performance study?	no
Other information	
Any other information considered useful	see PM 7/079 (2) for grapevine flavescence doree phytoplasma.
The following complementary files are available online:	
	<ul style="list-style-type: none"> • Validation report • Validation report annex • Validation report addendum • 2020.molbio.012 Detection and identification of Candidatus Phytoplasma fragariae V1.0

Creation date: 2025-06-30 09:30:41 - Last update: 2026-04-09 09:50:40