

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
(11-17239)

Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Target Organism	Meloidogyne	
Short description	Detection of Meloidogyne sp. by extraction with automated zonal centrifuge (AZC) from soil and potato tubers, followed by microscopy	
Laboratory contact details	ILVO Institute for Agricultural and Fisheries Research Burg. Van Gansberghelaan 96, 9820 Merelbeke - Melle, Belgium	
Date and reference of the validation report	2008 - F16_N01	
Validation process according to EPPO Standard PM 7/98:	Yes	
Reference of the test description	0 PM7/119 Nematode extraction :3.5 Centrifugal flotation: Automated zonal centrifugation PM7/41 Meloidogyne chitwoodi and Meloidogyne fallax	
Is the test the same as described in the EPPO DP?	Yes	
Is the lab accredited for this test?	Yes	
Plant species tested (if relevant)		
Matrices tested (if relevant)	soil (different types), potato tubers	
List of methods used		
Method for extraction / isolation / baiting of target organism from matrix	X	Extraction with automated zonal centrifuge (AZC)
Molecular methods, e.g. hybridization, PCR and real time PCR		
Serological methods: IF, ELISA, Direct Tissue Blot Immuno Assay		
Plating methods: selective isolation		
Bioassay methods: selective enrichment in host plants, baiting, plant test and grafting.		
Pathogenicity test		
Fingerprint methods: protein		

profiling, fatty acid profiling & DNA profiling		
Morphological and morphometrical methods intended for identification	X	microscopy
Biochemical methods: e.g. enzyme electrophoresis, protein profiling		
Other		
Analytical sensitivity (= limit of detection)		
What is smallest amount of target that can be detected reliably?	<5 targets: 30% detection; 5-9 targets: 70% detection; >10 targets: 100% detection	
Diagnostic sensitivity		
Proportion of infected/infested samples tested positive compared to results from the standard test , see appendix 2 of PM 7/98		
Specify the standard test		
Analytical specificity		
Specificity value		
Number of strains/populations of target organisms tested		
Number of non-target organisms tested		
Cross reacts with (specify the species)		
Diagnostic Specificity		
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test		
Specify the standard test		
Reproducibility		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	100% for all levels	
Repeatability		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	>5 targets: 100%	
Test performance study		
Test performance study?	No	
Include brief details of the test performance study and its output.It available, provide a link to published article/report		

<u>Other information</u>	
Any other information considered useful e.g. robustness, ease of performing the test, etc.	method valid for different soil types