

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
(11-17239)

Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Target Organism	'Candidatus Liberibacter solanacearum'	
Short description	Diagnostic protocol for the detection and identification of 'Candidatus Liberibacter solanacearum' in carrot seeds (DNA extraction method, real-time PCR and conventional PCR)	
Laboratory contact details	Council for Agricultural Research and Economics– Research Centre for Plant Protection and Certification Via Carlo Giuseppe Bertero, 22, 00156 Rome, Italy	
Date and reference of the validation report	2017-05-25 - Ilardi V. , V. Lumia, E. Di Nicola, M. Tavazza, 2018. Identification, intra and inter-laboratory validation of a diagnostic protocol for ‘Candidatus Liberibacter solanacearum’ in carrot seeds. European Journal of Plant Pathology https://doi.org/10.1007/s10658-018-01606-w	
Validation process according to EPPO Standard PM 7/98:	Yes	
Reference of the test description	0 Ilardi V. , V. Lumia, E. Di Nicola, M. Tavazza, 2018. Identification, intra and inter-laboratory validation of a diagnostic protocol for ‘Candidatus Liberibacter solanacearum’ in carrot seeds. European Journal of Plant Pathology https://doi.org/10.1007/s10658-018-01606-w	
Is the test the same as described in the EPPO DP?	Modified Modified - DNA extraction was from Ilardi et al. (2018) European Journal of Plant Pathology https://doi.org/10.1007/s10658-018-01606-w - real-time PCR was a modification of Li et al. (2009) Journal of Microbiological Methods 78:59–65 Not modified - conventional PCR as reported by Ravindran et al., (2011) Plant Disease 95.12: 1542-1546.	
Is the lab accredited for this test?	No	
Plant species tested (if relevant)	Daucus carota (carrot)	
Matrices tested (if relevant)	seed	
List of methods used		
Method for extraction / isolation / baiting of target organism from matrix	X	The International Seed Federation (ISF, 2016) recommends testing samples of 20 g of Apiaceae seeds divided into two sub-samples of 10 g each. The ISF DNA extraction protocol (2016) was used with some modification. Seeds were washed by shaking them for 30 min in 0.5% Triton X-100 and, after several rinses, they were left to soften in

		<p>water overnight. The seeds were crushed with a mechanical homogenizer in heavy plastic bags (Bioreba) in 1:10 (w/v) of a modified Trimethylammonium bromide (CTAB) buffer (2,5% CTAB, NaCl 1.4 M, Tris-HCl 1 M pH 8.0, EDTA 0.5 M, pH 8.0, PVP-40 1%, 30 mM ascorbic acid). 400 µg of RNase A was added to 500 µl of homogenate (corresponding to 50 seeds), and after incubation at 65 °C for 30 min, total genomic DNA was extracted using a DNeasy Plant Mini Kit (Qiagen, Germany) following the manufacturer's instructions. DNA was eluted in 100 µl of AE buffer provided by the kit.</p>
Molecular methods, e.g. hybridization, PCR and real time PCR	X	<p>- Real-time PCR modified from Li et al. (2009) Journal of Microbiological Methods 78:59–65. The primers and probes were 'Candidatus Liberibacter spp. specific HLB_r primer and HLB_p probe, Ca. L. solanacearum specific LsoF primer. Deviations from the reference: PCR reagents (Universal master mix II no UNG -applied biosystem), each primer and probe concentrations (400nM and 150nM, respectively), DNA (1 µl) reaction volume (15µl). Amplification condition: 1 cycle 95°C/10 min, 45 cycles 95°C/15sec and 60°C/60 sec.</p> <p>- conventional PCR as reported by Ravindran et al., (2011) Plant Disease 95.12: 1542-1546.</p>
Serological methods: IF, ELISA, Direct Tissue Blot Immuno Assay		
Plating methods: selective isolation		
Bioassay methods: selective enrichment in host plants, baiting, plant test and grafting.		
Pathogenicity test		
Fingerprint methods: protein profiling, fatty acid profiling & DNA profiling		
Morphological and morphometrical methods intended for identification		
Biochemical methods: e.g. enzyme electrophoresis, protein profiling		
Other		
Analytical sensitivity (= limit of detection)		
What is smallest amount of target that can be detected reliably?	<p>The limit of detection (LOD), calculated with the total DNA extract of CaLsol infected seeds, was of 10⁻² and 10⁻³ dilution for the conventional and real-time PCR, respectively. For the real-time PCR, the LOD was also evaluated with purified pTXZC18 diluted with water. Five copies of the target were detected with Ct values of 34.57 ± 0.428 in 100% of the experiments (24/24)</p>	

Diagnostic sensitivity	
Proportion of infected/infested samples tested positive compared to results from the standard test , see appendix 2 of PM 7/98	
Specify the standard test	Results from Test performance study with 11 laboratories: real-time PCR: 98.6% conventional PCR: 100%
Analytical specificity	
Specificity value	
Number of strains/populations of target organisms tested	<p>32 target organisms were tested:</p> <ol style="list-style-type: none"> 1. ISPAVE_VIb_1 'Berlicum' carrot seed infected by CaLsol haplotype E (Ilardi et al., 2016) 2. 4Flakkée 'Flakkée' CaLsol infected carrot seed 3. 5Maestro 'Maestro' CaLsol infected carrot seed 4. ISPAVE_VIb_6 'Nantese 3' carrot seed infected by CaLsol haplotype D (Ilardi et al., 2016) 5. 8Berlicum 'Berlicum2' CaLsol infected carrot seed 6. ISPAVE_VIb_9 'Berlicum 2' carrot seed infected by CaLsol haplotype E (Ilardi et al., 2016) 7. 10Berlicum 'Berlicum 2' CaLsol infected carrot seed 8. ISPAVE_VIb_11 'Falkkée' carrot seed infected by CaLsol haplotype D (Ilardi et al., 2016) 9. ISPAVE_VIb_15 'Mezza lunga nantese' carrot seed infected by CaLsol haplotype E (Ilardi et al., 2016) 10. ISPAVE_VIb_17 'Berlicum' carrot seed infected by CaLsol haplotype D (Ilardi et al., 2016) 11. C-AV 'Nantese migliorata 2' CaLsol infected carrot seed 12. C1 carrot seed infected by CaLsol haplotype D 13. C2 carrot seed infected by CaLsol haplotype D 14. C3 carrot seed infected by CaLsol haplotype D 15. C4 carrot seed infected by CaLsol haplotype D 16. C5 carrot seed infected by CaLsol haplotype E/D 17. C6 carrot seed infected by CaLsol haplotype D 18. P4 parsley seed infected by CaLsol haplotype D 19. P1 parsley seed infected by CaLsol haplotype E 20. P2 parsley seed infected by CaLsol haplotype E 21. P3 parsley seed infected by CaLsol haplotype E 22. 1P 'Riccio Verde' CaLsol infected parsley seed 23. 2P 'Riccio Verde' CaLsol infected parsley seed 24. 3P 'Comune' CaLsol infected parsley seed 25. 4P 'Comune 2 multifoglia' CaLsol infected parsley seed 26. 5P 'Gigante' CaLsol infected parsley seed 27. 6P 'Gigante' CaLsol infected parsley seed 28. 7P 'Gigante d'Italia' CaLsol infected parsley seed 29. P 1 SCS 'Gigante di Napoli' CaLsol infected parsley seed 30. P 4 SCS 'Gigante d'Italia' CaLsol infected parsley seed 31. P 5 SCS 'prezzemolo Comune 2' CaLsol infected parsley seed 32. P 7 SCS 'Riccio Muschiato' CaLsol infected parsley seed 33. S1 'Sedano D'Elne' CaLsol infected celery seed 34. S-AV 'Sedano D'Elne' CaLsol infected celery seed 35. CaLsol control pTXZC18 plasmid with the CaLsol 16S rDNA target (Li et al., 2009) kindly provided by Li 2009. <p>In the test performance study with 11 laboratories the</p>

	<p>following samples were tested:</p> <p>C4 carrot seed infected by CaLsol haplotype D, ISPAVE_Vlb_1 'Berlicum' carrot seed infected by CaLsol haplotype E, C-AV 'Nantese migliorata 2' CaLsol infected carrot seed, and for real-time PCR also CaLsol control pTXZC18 plasmid.</p>
Number of non-target organisms tested	<p>37 non-target organisms were tested:</p> <ol style="list-style-type: none"> 1. 12Nantese2 'Nantese2' CaLsol free carrot seed 2. 14Berlicum2 'Berlicum2' CaLsol free carrot seed 3. 16LungaB. 'Lunga di Berlicum' CaLsol free carrot seed 4. P 2 SCS 'Gigante di Napoli' CaLsol free parsley seed 5. P 3 SCS 'Gigante di Napoli' CaLsol free parsley seed 6. P 6 SCS 'Comune 2' CaLsol free parsley seed 7. P-AV 'Gigante di Napoli' CaLsol free parsley seed 8. 2Berlicum 'Berlicum' CaLsol free carrot seed 9. 3Bolero 'Bolero F1' CaLsol free carrot seed 10. 7Nantes3 'Nantes3' CaLsol free carrot seed 11. 13Nantese2 'Nantese2' CaLsol free carrot seed 12. S2 'Peros Rendy' CaLsol free celery seed 13. S3 'Sedano D'Elne' CaLsol free celery seed 14. F1 'Montebianco' CaLsol free fennel seed 15. F3 'Wadenromen' CaLsol free fennel seed 16. F4 'Romanesco' sel. Circeo CaLsol free fennel seed 17. F-AV 'Wadenromen' CaLsol free fennel seed 18. 1519 <i>Pseudomonas fluorescens</i> 19. 1174 <i>P. putida</i> 20. 1182 <i>P. marginalis</i> from chicory 21. 1146 <i>P. syringae</i> pv <i>syringae</i> from lemon 22. 1001 <i>Agrobacterium tumefaciens</i> 23. 1235 <i>Erwinia herbicola</i> ISF438 24. 1030 <i>Xantomonas campestris</i> pv <i>campestris</i> from cabbage 25. 1049 <i>Xantomonas arboricola</i> pv <i>corylina</i> from turnip 26. 1240 <i>Pectobacterium carotovora</i> from artichoke 27. 1433 <i>Pectobacterium carotovora</i> from zucchini 28. 04-500 <i>X. campestris</i> pv <i>begoniae</i> from carrot 29. 11-267N2 <i>Pseudomonas</i> sp from fennel 30. 1432 <i>P. viridiflava</i> from tomato 31. Ferr1 <i>Phytoplasma stolbur</i> (solani 16SrXII-A) 32. PAV 1 Unknown bacterium from carrot seed 33. PAV 2 Unknown bacterium from carrot seed 34. PAV 3 Unknown bacterium from carrot seed 35. PAV 4 Unknown bacterium from carrot seed 36. PAV 5 Unknown bacterium from carrot seed 37. PAV 6 Unknown bacterium from carrot seed <p>In the test performance study with 11 laboratories the following samples were tested: F-AV 'Wadenromen' CaLsol free fennel seed, F1 'Montebianco' CaLsol free fennel seed, 3Bolero 'Bolero F1' CaLsol free carrot seed, 04-500 <i>X. campestris</i> pv <i>begoniae</i> from carrot, 11-267N2 <i>Pseudomonas</i> sp from fennel)</p>
Cross reacts with (specify the species)	None of them
Diagnostic Specificity	
Proportion of uninfected/uninfested	real-time PCR: 100%

samples (true negatives) testing negative compared to results from a standard test	conventional PCR: 81.5%
Specify the standard test	
<u>Reproducibility</u>	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	real-time PCR: 100% conventional PCR: 100% calculated with the total DNA extract of CaLsol infected seeds at 10 ⁻² and 10 ⁻³ dilution for the conventional and real-time PCR, respectively. For the real-time PCR, was also evaluated with Five copies of purified pTXZC18 diluted with water.
<u>Repeatability</u>	
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	real-time PCR: 100% conventional PCR: 100% 2 different operators for real-time PCR and 3 for conventional PCR. 2 different equipments for real-time PCR and 2 for conventional PCR. Calculated with the total DNA extract of CaLsol infected seeds at 10 ⁻² and 10 ⁻³ dilution for the conventional and real-time PCR, respectively. For the real-time PCR, was also evaluated with Five copies of purified pTXZC18 diluted with water.
<u>Test performance study</u>	
Test performance study?	Yes
Include brief details of the test performance study and its output. If available, provide a link to published article/report	<p>Ilardi V. , V. Lumia, E. Di Nicola, M. Tavazza, 2018. Identification, intra and inter-laboratory validation of a diagnostic protocol for 'Candidatus Liberibacter solanacearum' in carrot seeds. European Journal of Plant Pathology https://doi.org/10.1007/s10658-018-01606-w</p> <p>TPS was performed by ten laboratories of the Italian Regional Plant Protection Service (IRPPS), widespread throughout the country, and CREA-DC-Rome laboratory. For real-time and conventional PCR, the oligonucleotides and reagents, including water, were sent to the participants. For the end-point PCR and real-time PCR, Go Taq G2Flexi DNA polymerase (Promega) and TaqMan Universal Master Mix II (applied biosystem) were provided, respectively. Each sample was tested by the participants in triplicate (technical replicates). To test the DNA extraction protocol, CaLsol infected and CaLsol free seeds, were provided together with the buffers and the DNeasy Plant Mini Kit (Qiagen, Germany).</p> <p>real-time PCR Li et al., 2009 Diagnostic sensitivity 98.6% Diagnostic specificity 100.0% Relative accuracy 99.0% Accordance 98.2% Concordance 98.0% COR* 1.11</p>

	<p>end-point PCR Ravindran et al., 2011 Diagnostic sensitivity 100% Diagnostic specificity 81.5% Relative accuracy 88.9% Accordance 82.2% Concordance 80.0% COR* 1.15</p> <p>seed DNA extract evaluated by real-time PCR Li et al., 2009 Diagnostic sensitivity 100.0% Diagnostic specificity 95.0% Relative accuracy 98.75% Accordance 97.81% Concordance 97.5% COR* 1.14</p> <p>seed DNA extract evaluated by end-point PCR Ravindran et al., 2011 Diagnostic sensitivity 90.74% Diagnostic specificity 100.0% Relative accuracy 93.82% Accordance 90.12% Concordance 88.20% COR* 1.22</p> <p>*Concordance odds ratio= $\text{accordance} \times (100 - \text{concordance}) / (\text{concordance} \times (100 - \text{accordance}))$, to address the variability of the method within and between laboratories, calculated as indicated by ISO 16140:2003.</p>
<u>Other information</u>	
Any other information considered useful e.g. robustness, ease of performing the test, etc.	