

EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION
ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES
(11-17239)

Summary sheet of validation data for a diagnostic test

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

Target Organism	Meloidogyne chitwoodi Meloidogyne fallax	
Short description	Identification of Meloidogyne chitwoodi and M. fallax by SCAR PCR	
Laboratory contact details	Anses, Laboratoire de la Santé des Végétaux - Unité de Nématologie Domaine de la Motte au Viconte BP 35327, 35653 Le Rheu, France	
Date and reference of the validation report	september 2010 - validation report september 2010	
Validation process according to EPPO Standard PM 7/98:	Yes	
Reference of the test description	0 PM7/41 (2) - Appendix 3 Zijlstra (2000) Identification of Meloidogyne chitwoodi, M. fallax and M. hapla based on a SCAR PCR : a powerful way of enabling reliable identification of populations on individuals that share common traits. European Journal of Plant Pathology 106, 283-290	
Is the test the same as described in the EPPO DP?	Yes	
Is the lab accredited for this test?	Yes	
Plant species tested (if relevant)		
Matrices tested (if relevant)	isolated nematodes	
List of methods used		
Method for extraction / isolation / baiting of target organism from matrix		
Molecular methods, e.g. hybridization, PCR and real time PCR	X	species specific SCAR PCR (multiplex PCR; i.e one primer set for each species)
Serological methods: IF, ELISA, Direct Tissue Blot Immuno Assay		
Plating methods: selective isolation		
Bioassay methods: selective enrichment in host plants, baiting, plant test and grafting.		
Pathogenicity test		

Fingerprint methods: protein profiling, fatty acid profiling & DNA profiling		
Morphological and morphometrical methods intended for identification		
Biochemical methods: e.g. enzyme electrophoresis, protein profiling		
Other		
<u>Analytical sensitivity (= limit of detection)</u>		
What is smallest amount of target that can be detected reliably?	1 J2 for M. fallax and 2 J2 for M. chitwoodi	
<u>Diagnostic sensitivity</u>		
Proportion of infected/infested samples tested positive compared to results from the standard test , see appendix 2 of PM 7/98		
Specify the standard test		
<u>Analytical specificity</u>		
Specificity value	100% for M. fallax and 100% for M. chitwoodi	
Number of strains/populations of target organisms tested	2 for M. fallax and 4 for M. chitwoodi (for details see annex 1 from validation report)	
Number of non-target organisms tested	31 nematodes populations (for details see neex1 from validation report)	
Cross reacts with (specify the species)	none	
<u>Diagnostic Specificity</u>		
Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test		
Specify the standard test		
<u>Reproducibility</u>		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	91% (1 J2), 100% (2 J2) for M. fallax 83% (2 J2), 100% (5 J2) for M. chitwoodi	
<u>Repeatability</u>		
Provide the calculated % of agreement for a given level of the pest (see PM 7/98)	75% (1 J2), 100% (2 J2) for M. fallax 62% (2 J2), 100% (5 J2) for M. chitwoodi	
<u>Test performance study</u>		
Test performance study?	No	
Include brief details of the test performance study and its output.It available, provide a link to published article/report		

<u>Other information</u>	
Any other information considered useful e.g. robustness, ease of performing the test, etc.	The full report is available upon request to the laboratory.